

**INDUSTRIAL TECHNOLOGY DEVELOPMENT INSTITUTE**  
(Formerly National Institute of Science and Technology)

STANDARDS AND TESTING DIVISION  
Gen. Santos Ave., Bicutan, Taguig, Metro Manila 1631

Fax No.: (632) 837-31-67 / 837-00-32

Tel. Nos. 837-20-71 to 82  
local 2188, 2189

**REPORT OF ANALYSIS**  
**No. 0308ICL0028**


Customer's Name : Gtech Dental Laboratory  
Address : 57 Sampaguita St., sampaguita Village, Malanday, Marikina City  
1805 Philippines

Sample : *All Ceramic Crown Using IPS EMAX*  
Description : 6 pcs  
Identification :  
Date Received : March 28, 2008  
Date(s) Tested : April 1 - 4, 2008


Extractable Lead (Pb), mg/L Not Detected

Method Detection Limit: Pb=0.05mg/L  
Test Method: ASTM C738-94

**VALIDITY OF THE REPORT:** The test results are those obtained at the time of the test and pertain only to the sample(s) received by this Laboratory. *Codes and words in italics* are quoted solely for the customer's reference; significance of these codes and words are not verified by this Laboratory. This report is not to be used for advertising purposes or sales promotion.

  
**CONCEPCION P. GAYOMALI**  
Head, Inorganic Chemistry Laboratory

Issued under the Authority of:

  
**HERMELINA H. BION**  
Division Chief