

**INDUSTRIAL TECHNOLOGY DEVELOPMENT INSTITUTE**  
(Formerly National Institute of Science and Technology)

STANDARDS AND TESTING DIVISION  
Gen. Santos Ave., Bicutan, Taguig, Metro Manila 1631

Fax No.: (632) 837-31-67 / 837-00-32

Tel. Nos. 837-20-71 to 82  
local 2188, 2189

**REPORT OF ANALYSIS**  
**No. 0308ICL0023**

Customer's Name : GTECH Dental Laboratory  
Address : # 57 Sampaguita St., Sampaguita Village, Malanday, Marikina City


Sample : *Metal Coping Using Talladium Alloy (non-precious)*  
*Molar Metal Coping*

Description : 6 pcs  
identification :  
Date Received : March 13, 2008  
Date(s) Tested : March 24 - 26, 2008


Extractable Lead (Pb), mg/L      Not Detected

Method Detection Limit: Pb=0.05mg/L  
Test Method: ASTM C738-94

**VALIDITY OF THE REPORT:** The test results are those obtained at the time of the test and pertain only to the sample(s) received by this Laboratory. *Codes and words in italics* are quoted solely for the customer's reference; significance of these codes and words are not verified by this Laboratory. This report is not to be used for advertising purposes or sales promotion.

  
JOSEFINA T. HERNANDEZ  
OIC, Inorganic Chemistry Laboratory

Issued under the Authority of:

  
HERMELINA H. BION  
Division Chief