

Republic of the Philippines
Department of Science and Technology

INDUSTRIAL TECHNOLOGY DEVELOPMENT INSTITUTE
(Formerly National Institute of Science and Technology)

STANDARDS AND TESTING DIVISION
Gen. Santos Ave., Bicutan, Taguig, Metro Manila 1631

Fax No.: (632) 837-31-67 / 837-00-32

Tel. Nos. 837-20-71 to 82
local 2188, 2189

REPORT OF ANALYSIS
No. 0308ICL0027


Customer's Name : Gtech Dental Laboratory
Address : 57 Sampaguita St., sampaguita Village, Malanday, Marikina City
1805 Philippines

Sample : *Removable Denture*
Description : 6 pcs
Identification : *Made of chrome cobalt metal frame & acrylic*
Date Received : March 28, 2008
Date(s) Tested : April 1 - 4, 2008


Extractable Lead (Pb), mg/L Not Detected

Method Detection Limit: Pb=0.05mg/L
Test Method: ASTM C738-94

VALIDITY OF THE REPORT: The test results are those obtained at the time of the test and pertain only to the sample(s) received by this Laboratory. *Codes and words in italics* are quoted solely for the customer's reference; significance of these codes and words are not verified by this Laboratory. This report is not to be used for advertising purposes or sales promotion.


CONCEPCION P. GAYOMALI
Head, Inorganic Chemistry Laboratory

Issued under the Authority of:


HERMELINA H. BION
Division Chief